

DEPARTMENT OF CHEMISTRY

UNIVERSITY OF VIRGINIA

SAFETY MANUAL

EMERGENCY PHONE NUMBERS

ALL EMERGENCIES	911
UNIVERSITY POLICE	4-7166
ENVIRONMENTAL HEALTH & SAFETY	2-4911
http://www.ehs.virginia.edu/	

(REVISED: September 2011)

**ACCIDENT/INJURY
on the job?**

MAJOR: CALL 911 and Follow Instructions

MINOR: GET TREATMENT IMMEDIATELY at one of the following:

Daytime: UVA- WorkMed --- 243-0075
545 Ray C. Hunt Drive, UVA Fontaine Research Park

UVA Employee Health --- 924-2013
1222 Jefferson Park Avenue

Dr. W. R. Dandridge --- 977-6622
1149 Rose Hill Drive

Med Express --- 978-3998
1149 Seminole Trail

Nights/Weekends: UVA Emergency Room --- 924-2231
Lee Street

Martha Jefferson Emergency Room ---654-7150
500 Martha Jefferson Drive

Please also contact your supervisor and department chair's office as soon as possible with details and treatment paperwork.

FORWARD

1. This manual provides a summary of the Chemistry Department safety policies and standards. Read it carefully and direct your unanswered safety questions to your advisor or any of the safety committee members listed below.
2. After reading this safety manual, **complete, sign, and date the Certification Form at the end of this manual**. Give the completed form to the Susan Marshall (Rm. #188). You are not allowed to use any of the Chemistry laboratories prior to turning in this form.
3. You are responsible for compliance with all safety regulations and for elimination of hazards in your own lab. It is your responsibility to make your lab a safe place to work for you, your lab partners and visitors to your lab. Safe work habits which you develop now will prepare you for work in industry and may save you from injury or save your life.
4. You must also read carefully the University of Virginia Laboratory Survival Manual and attend the mandatory safety training session at the earliest offering.

Office

Home/Cell

Chemistry Safety Committee

Robert Bryant - Chair	924-1494	823-2454
Ralph Allen	924-3577	434-531-4250 (cell)
Charles Grisham	924-7012	977-1572
Cameron Mura	924-7824	434-249-3035
Jason Chroma	243-2131	434-326-0336

Department Chair

W. Dean Harman	924-3060
----------------	----------

ACKNOWLEDGEMENT

This manual has been developed in part using that created by G. Carta, Chemical Engineering Department and based upon a laboratory safety manual used at the University of Delaware. We are grateful to Professor Carta and Mr. George Whitmyre of the Chemical Engineering Department at the University of Delaware for his permission to use and modify parts of their manual.

TABLE OF CONTENTS

	Page
Chapter 1. General Safety Principles and Regulations	5
Where to Get Advice on Safety Problems	10
Safety Inspections and Enforcement	10
Chapter 2. Assembly and Use of Apparatus	11
Chapter 3. Handling and Storage of Chemicals and Solvents	14
Chapter 4. Compressed Gases and Gas Regulators	20
Chapter 5. Tool and Machine Safety	22
Appendix 1. Eye Protection	23
Appendix 2. Waste Disposal Guidelines	24
Appendix 2A. Waste Disposal Labels	27
Appendix 3. Hazard Review Checklist	29
Appendix 4. References	34
Emergency Notification Form	35
Certification form	36

CHAPTER 1

GENERAL SAFETY PRINCIPLES AND REGULATIONS

Know the safety procedures that apply to the work being done. Determine the potential hazards (e.g., physical, chemical biological, radiological) and appropriate safety precautions before beginning any new operation. Every laboratory worker should observe the following rules:

1. **Wear safety glasses or goggles** at all times in designated areas (see Appendix 1 for details). Contact lenses are permitted in these areas provided that safety glasses are also worn.
2. **Avoid carrying out experimental work alone in a laboratory building.** Under ordinary circumstances, someone else should be present in your laboratory area in order to render emergency help should this be required. When working at night or on weekends, make sure that someone else is notified of your presence.
3. **All injuries** (and near misses), no matter how slight, and accidents of any kind should be reported to a member of the Safety Committee or the Department Chairman without delay.
4. **Shoes** with uppers of a solid material must be worn in the laboratories. This will prevent permeation by liquid chemical spills and other injuries. Open-toed shoes, flip-flops, and bare feet are not permitted.
5. **Loose** (e.g. saris, dangling scarves or neckties, overlarge coats), **skimpy** (e.g. shorts, halter tops), or **torn clothing** should not be worn in laboratory areas. Loose or torn clothing can easily catch fire, dip into chemicals, knock items off benches, or become ensnared in apparatus or moving machinery. Skimpy clothing offers little protection to the skin in the event of chemical splash.
6. **Avoid Unrestrained long hair or long beards** in laboratory areas. These can easily catch fire, dip into chemicals, or become ensnared in apparatus or moving machinery.
7. **Clearly label all chemicals** and keep them in capped containers. Parafilm and aluminum foil are not acceptable for long-term storage of chemicals. Screw caps should be used whenever possible. Post warning signs when unusual hazards, such as radiation, flammable materials, biological hazards, or other special problems exist. Place your initials and date on the label of any chemical container.
8. Properly **dispose of laboratory waste** and unwanted chemicals by following strictly the procedures outlined in Appendix 2. All HAZARDOUS WASTE DISPOSAL CONTAINERS MUST BE PROPERLY LABELED AND KEPT CAPPED AT ALL TIMES ACCORDING TO THE PROCEDURES IN APPENDIX 2.
9. All **departing graduate and undergraduate students, staff, post-docs, visitors, and faculty,** must complete and sign the exit form available from the department office

ATTESTING THAT ALL CHEMICAL WASTES GENERATED HAVE BEEN PROPERLY DISPOSED AND THAT THE LABORATORY AREAS USED HAVE BEEN LEFT IN GOOD ORDER. The form must be signed by the faculty advisor. Completion and approval of the form is required prior to receiving a degree or final salary compensation.

10. **Promptly and completely neutralize, absorb, and bag spilled materials by the safest means possible.** The Environmental Health & Safety Office is equipped for a rapid response for toxic, flammable or radioactive material spills; if you are uncertain about the nature of the spill or on how to take care of it, call the Environmental Office. They may authorize you to perform a spill cleanup if the hazards are insignificant. Otherwise, the Environmental Office will undertake the cleanup operation.

Promptly bag and dispose of oily or solvent-saturated cleanup materials following the instructions for waste disposal given below. The Environmental Health and Safety Office can determine if special containment and labeling is needed.

11. **Mercury spills:** Call 2-4911 for help and equipment to clean up mercury spills.

All instruments containing mercury are potential sources of poisonous mercury vapor and require special attention. Whenever possible, these devices should have a catch tray to contain spills. Whenever possible, mercury thermometers should be replaced with non-mercury ones. The Environmental Health and Safety Office has offered to provide replacement thermometers at no charge.

12. Contamination of food, drink and smoking materials is a potential route for exposure to toxic substances. **Consumption of food and beverages** should be reserved to office rooms and desks not in chemical laboratory spaces. **Smoking is not permitted.**

13. **Glassware or utensils** that have been used for laboratory operations should never be used to prepare or consume food or beverages. Laboratory refrigerators, ice machines, ice chests and such should not be used for food storage.

14. **Wash well before leaving the laboratory area.** However, avoid the use of solvents for washing the skin. They remove the natural protective oils from the skin and can cause irritation and inflammation. In some cases, washing with a solvent may facilitate adsorption of a toxic chemical.

15. All laboratories must display updated **lab occupants information** with contact information to be used in case of emergencies. A sample format is given at the end of this Chapter.

16. **Use equipment only for its designed purpose.** The use of makeshift tools and shortcut methods leads to equipment damage and injuries. If you are in doubt, seek the help of your advisor or the Safety Committee. Label broken equipment and notify your advisor or lab coordinator promptly and provide for repairs and/or storage.

17. Familiarize yourself with **emergency procedures** and learn how to obtain additional help in any emergency. Know how to use the emergency equipment in your work area. **EVERYONE MUST BE FAMILIAR WITH THE LOCATION AND USE OF SAFETY SHOWERS, EYE WASH EQUIPMENT AND PERSONAL PROTECTIVE EQUIPMENT.** Take a moment to familiarize yourself with the locations of these important items.
18. **Flush eye wash stations periodically.** Eye wash stations in your laboratory area should be flushed periodically, preferably on weekly basis. Safety showers are inspected by UVA Facilities Management personnel with the inspection date recorded and attached to the shower. Make sure the inspections are conducted on a timely basis by checking the tag in the safety shower in your area. Should the date be more than 1 year old, call Facilities Management (924-1777) and request an inspection.
19. Discharged **fire extinguishers** must be recharged and returned to service immediately.
20. Do not cover **windows of laboratory doors** except for special experimental requirements, as passers-by should have an unobstructed view to notice if someone is in distress and needs help.
21. **All laboratory and graduate student entry doors must be locked at all times when not occupied.** Access doors should always be locked after 5 p.m. and on weekends.
22. **Doors between adjacent laboratories must never be blocked.** These will provide alternate escape routes in case of an emergency.
23. **No unaccompanied visitors are allowed in any laboratory or graduate student office at any time.** Please, make sure that your visitors follow all safety rules listed in this manual. It is your responsibility to provide them with proper personal safety equipment.
24. **Pets of any kind are not to be brought into the laboratories.** Pets are likely to upset equipment or be poisoned by ingesting or contacting toxic chemicals that may be present.
25. **Hypodermic needles or GC/HPLC syringes** require special attention. Safe lab practice requires that sharp objects, especially needles and syringes, be protected to avoid accidental injection into the skin. A GC septum or a plastic syringe cap will protect these points adequately.
 - Needle disposal boxes labeled “bio-hazard” must be used for needles and syringes used with biological samples where bio-hazards can exist.
 - Needle disposal boxes for “non-bio-hazard” needles should be used for other needles, and are available from the Environmental Health and Safety Office.
26. To reduce fire hazards and general clutter, empty cardboard boxes, crates and solvent bottles must be **removed from laboratory areas within 24 hours after receipt.**

27. Compliance with the Virginia Occupational Safety and Health Hazard Communication Standard, requires that **Material Safety Data Sheets (MSDS)** be made available to all University employees. These sheets provide detailed information and precautionary measures for the handling of chemicals and solvents. The MSDS are available online at ehs.virginia.edu/ehs/ehs.chemicalsafety.html.
28. An unintentional cross connection between the University water supply and waste water may occur from a **submerged inlet** in your laboratory unless vacuum breakers are present on the faucet. "Draw down" occurs when city water pressure drops from low reservoir conditions, opening of fire hydrants, unbalanced demand on water circuits or from other causes. If your lab faucets have an attached Tygon or rubber hose to prevent splashing or to facilitate washing, it may siphon sink wastes and possibly raw sewage into the water lines. You must cut off each faucet hose at least 2 inches (5 cm) above the sink rim elevation. This gap will assure you that no laboratory sewage will siphon into our potable water supply.

A common problem is **dry drain traps**". A plumbing trap which has lost water through evaporation is likely to release sewer gases into your laboratory. Dry traps in adjacent rooms may also duct lab odors from remote labs into yours; this may account for unlocatable mystery odors that may be noticed. Pouring water into seldom-used drains bi-weekly will restore the water seal and assure you that no sewer gases will escape into your work environment. Remember floor drains as well.
29. **Be alert to unsafe conditions and actions** and call attention to them so that corrections can be made as soon as possible. Someone else's accident can be as dangerous to you as any you might have yourself.
30. **Avoid distracting or startling any other worker.** Practical jokes or horseplay cannot be tolerated at any time.
31. Loud radios are prohibited, and head phones not permitted because you could not hear warnings or calls for help.
32. **Think, act and encourage safety until it becomes a habit.**

Chemistry Laboratory Contact Information

Name

Phone

Laboratory Name

xxx-xxxx

Advisor

Advisor Name

xxx-xxxx (office)
xxx-xxxx (home)

Students

Student Name

xxx-xxxx

Student Name

xxx-xxxx

Student Name

xxx-xxxx

Student Name

xxx-xxxx

Student Name

xxx-xxxx

Last updated: _____

WHERE TO GET ADVICE ON SAFETY PROBLEMS

When you become aware of a safety problem, your advisor, the Safety Committee, and the Office of Environmental Health and Safety (EHS) personnel are the best source for information. **Day-to-day problems should be directed to your advisor or lab coordinator.** In the case of equipment design problems and new equipment setups, the Hazard Review Checklist, given in Appendix 3 is helpful.

The Office of Environmental Health and Safety (EHS, Phone: 2-4911, <http://www.ehs.virginia.edu/>) provides trained safety professionals for administration and implementation of safety at the University. They carry out waste disposal, fume hood tests, etc. In addition they are responsible for radiation safety, fire safety, and other major hazard control areas.

SAFETY INSPECTIONS AND ENFORCEMENT

Periodic safety inspections are conducted in the Department labs to check compliance with our safety regulations. The inspection team is composed of the Safety Committee and graduate student representatives. Results of the inspection are summarized in a report directed to the Department Chairman, with copies to everyone in the Chemistry Department.

These periodic inspection reports will help you identify safety hazards in your lab and will remind you of our routine safety requirements. **Your safety is your primary responsibility. Equipment and procedures will be shut down if they are not in accordance with the established Departmental and University safety practices.**

The safety know-how and training you acquire in your work will prepare you for work in industry or in other organizations. Safety training at the University will enhance your value to an employer. A poor safety record can be a serious impediment to employment.

Faculty advisors or their designated group supervisor are responsible for routine auditing of their assigned laboratory, undergraduates, graduate students and post-doctoral fellows.

CHAPTER 2

ASSEMBLY AND USE OF APPARATUS

Prior to equipment set-up, review the Hazard Review Checklist (Appendix 3) to determine if adequate safety plans were considered in your equipment design.

The following are useful guidelines.

1. **Any equipment or experiment that is operated unattended** for any length of time must have emergency information displayed clearly.
2. **The apparatus should be set up in a clean and dry area.** Be certain that the equipment is firmly clamped and is kept well back from the edge of the laboratory bench. Many accidents occur when someone walks by a bench and brushes against the glassware or other equipment. Make sure that you use the proper size equipment for the experiment, allowing at least 20% free space. Flasks that contain solutions to be refluxed should have 50% free space. Position and clamp reaction apparatus thoughtfully in order to permit manipulation without the need to move the apparatus until the entire reaction is completed. Combine reagents in appropriate order, and avoid adding solids to hot liquids.
3. **Never use glassware that is chipped, cracked, etched or flawed in any way.**
4. **Keep work space uncluttered.** Only the required materials, instructions, notebook and pen should be present. Keep the work area free from extraneous chemicals, scraps of paper and paper towels. Keep all other equipment far back where it will not be knocked over.
5. **Ground glass joints or stopcocks should be sleeved with Teflon or freshly lubricated** unless a lubricant will contaminate the system. Retainer rings should be used on stopcock plugs.
6. **Condensers** must be properly supported with securely positioned clamps. Any attached water hoses should be clamped with clamps of an adequate material. Condensers running unattended overnight should preferably be attached to a water pressure regulator in order that surges in the water pressure do not cause the hoses to rupture.
7. **Overhead Stirrer motors** should be secured to retain proper alignment. An air driven stirrer or magnetic stirrer should be used whenever possible. Only non-sparking motors ought to be used in hazardous areas where significant amounts of flammable gases and solvents are present.
8. **A common injury** sustained in the laboratory occurs from the **improper insertion of glass tubing into a rubber stopper**. To avoid injuries while cutting glass tubing, hold the tubing against a firm notched support, make one quick firm stroke with a sharp file, rocking the file to extend the deep nick one-third around the circumference. Hold the tubing in both hands,

away from the body, with the nick turned directly opposite the body. Place the thumbs on the tubing opposite the nick about an inch apart. With hand protection, push out on the tubing with the thumbs. All glass tubing and rods should be fire polished before use.

When inserting glass tubing into a stopper, use a glove or towel for protection and be certain that the tubing is lubricated lightly and that excess pressure is not applied to the tubing.

9. **BROKEN GLASS, INCLUDING BROKEN REAGENT BOTTLES, SHOULD BE DISPOSED OF PROMPTLY ONLY IN SPECIALLY DESIGNATED CARDBOARD CONTAINERS.**

UNBROKEN chemical reagent, salt, and solvent bottles can be discarded in trash bins. Before discarding in a trash can, these containers must be THOROUGHLY RINSED. If a chemical label is present, it should be removed or defaced.

10. **Vacuum pumps and other belt-driven equipment** must always have a belt guard.
11. If a **cooling bath** is required and ice water is not cold enough, dry ice in an organic liquid should be used instead of liquid nitrogen whenever possible. The ideal cooling liquid for a dry ice bath should be relatively non-toxic, non-viscous, non-flammable, non-volatile, insoluble in water and should float dry ice. Ethylene glycol thinned with 2/3 water or isopropanol makes a useful cooling mixture.
12. **Hardware, regulators, glassware, solvents, dry chemicals, acids, etc., stored in the laboratory must be isolated from each other** in separate storage areas to prevent breakage and to avoid other undesirable effects.
13. **Electrical equipment including variacs, stirrers, vacuum pumps, etc., must be carefully checked for faulty or frayed line cords.** Grounded electrical plugs should be used: existing ungrounded plugs should be changed immediately.

EMERGENCY NOTIFICATION FORM

DATE: _____

TITLE OF EXPERIMENT: _____

RESEARCHER: _____

ADVISOR: _____

WARNINGS:

SPECIAL EMERGENCY PROCEDURES:

IN CASE OF EMERGENCY CALL:

1) _____ Ph. No.

2) _____ Ph. No.

3) _____ Ph. No.

POTENTIAL HAZARDS: (Toxic gases, Flammable solvents, Flammable gases, High pressure gas, Biological hazard, Radiation Hazard, etc.)

CHAPTER 3

HANDLING AND STORAGE OF CHEMICALS AND SOLVENTS

The following general guidelines apply to the handling and storage of chemicals and solvents.

1. Wear eye protection when performing any chemical operation. Protect your skin from chemical exposure. Gloves made of the appropriate material are usually advised.
2. A **fume hood** or other approved ventilation/exhaust system should be used whenever flammable solvents or toxic gases are used or generated. Remember the best ventilating protection results with the hood sash closed. Keeping all items 6 inches behind the sash line and minimize the quantity of equipment within the hood area improves its safety.

The operating condition of a hood should be determined before the hood is put to use. Fume hoods must be **inspected and certified annually by the Office of Environmental Health & Safety**. The Office personnel will affix a label on the fume hood indicating the operability of the system and the maximum acceptable elevation of the hood sash. If your fume hood has not been inspected in more than one year, please call EH&S.

Do not operate the fume hood with the sash above the indicated level.

3. All chemicals must be organized and stored on shelves or in cabinets where they will not be knocked over. One way to organize chemicals is to store organics by number of Carbon atoms and keep them separate from inorganics. Upon receipt, date and initial the label so that the age of the stock can be determined. Store chemicals in chemically compatible groups. Simple alphabetical storage may be dangerous!
4. **Flammable solvents:**
 - A. Properties of flammable liquids:
 1. **Flash Point:** Temperature at which the vapor pressure is sufficient to form an ignitable mixture with the air.
 2. **Ignition Temperature:** Minimum temperature required to cause self-sustained combustion.
 - B. **Classification of flammable liquids:**
 - Class IA Liquids: flash point below 73°F and boiling point below 100°F.
 - Class IB Liquids: flash point below 73°F and boiling point at or above 100°F.
 - Class IC Liquids: flash point between 73°F and 100 F.
 - Class II Liquids: flash point between 100°F and 140°F.
 - Class IIIA Liquids: flash point between 140°F and 200°F.
 - Class IIIB Liquids: flash point above 200°F.

A table of common solvents is given at the end of this chapter.

C. The **maximum allowable size of containers for flammable liquids** is as follows:

Container	Class: IA	IB	IC	II	III
Glass/plastic	1 pt	1 gal	1 gal	1 gal	1gal
Tinplate can	1 gal	5 gal	5 gal	5 gal	5 gal
Safety cans	2 gal	5 gal	5 gal	5 gal	5 gal

5. **Oxidizing agents**

A. **Peroxides, hydroperoxides, and peroxyesters** - these compounds are all active oxygen-containing materials which can decompose generating oxygen or oxidizing agents. These materials are chemically unstable to varying degrees.

Organic peroxides are among the most hazardous chemicals handled in a laboratory. Many organic compounds, including the following types, are known to form extremely dangerous peroxides.

1. Ethers, especially cyclic ethers such as THF.
2. Compounds containing benzylic hydrogen atoms, e.g. cumene.
3. Compounds containing the allylene ($\text{CH}_2=\text{CHCH}_2\text{R}$ structure).
4. Ketones.
5. Vinyl and vinylidene compounds; e.g., vinyl acetate and vinylidene chloride.

Examples of common materials which form dangerous peroxides upon long exposure to air are: Cyclohexene, Cyclooctene, Decalin, p-Dioxane, Ethyl ether, Isopropyl ether, Tetrahydrofuran (THF) and Tetralin.

B. **Disposal of Peroxides** - Do not mix with other chemicals for disposal – keep in a separate container properly labeled for disposal by the Environmental Office.

6. **Containers of Concentrated acids and bases should preferably be stored in non-metal (PTFC) trays**, separated from all other chemicals. They should not be stored on high shelves. **ACIDS AND BASES SHOULD BE STORED IN SEPARATE CABINETS!**

7. **All chemicals in the laboratory must be labeled with permanent labels**. The label should indicate the full chemical name or notebook number and the primary hazard associated with the substance (e.g., flammable, toxic). Do not use abbreviations. Include your initials and date. Self sticking labels fall off over time. Secure them with transparent tape.

8. **Only "non-hazardous materials" as defined in Appendix 2, may be poured into a sink.**
For all other materials follow the waste disposal guidelines in Appendix 2. Concentrated acids and bases can be removed by the Environmental Health and Safety Office. If you are unsure about any item, contact the EHS Office before putting anything in the sink.
9. **UNBROKEN** chemical reagent, salt, and solvent bottles can be discarded in trash bins. Before discarding in a trash can, these containers must be **THOROUGHLY RINSED**. If a chemical label is present, it should be removed or defaced.
10. **Dichromate in Sulfuric acid** and other strong acid or oxidizer cleaning solutions should not be used for general cleaning purposes. Due to liberation of extremely toxic chromyl chlorides, Dichromate/Sulfuric acid is approved for use only in fume hoods.
11. **Chemical toxins:**
 - A. **Cyanides and Azides** - are among the most toxic substances encountered in the Chemical laboratory. The compounds are toxic if inhaled, ingested or absorbed through the skin.
 - B. The **toxicity of common solvents** should be recognized. Solvents requiring special care include:
 1. Certain aromatic hydrocarbons such as benzene.
 2. Halogenated hydrocarbons
 3. Lower alcohols - methanol, ethanol, etc.
 - C. Do not overlook the **toxicity of chemical compounds**. It is best to consider every chemical toxic and to protect yourself accordingly.
12. **Shipment of hazardous materials originating from UVa**

There are critical safety and regulatory considerations for shipping materials, such as chemical and biochemical samples, from UVa. Particularly problematic are **hazardous materials**, including **dry ice** used to preserve biological samples. Discuss your shipping procedures with EHS before shipping anything.



To: All Departmental Chairs

From: Ralph Allen, Director

August 11, 2008

Shipment of Dangerous Goods (Hazardous Materials)

Recently Federal and State regulatory agencies have identified two instances where hazardous materials (dangerous goods) were illegally offered for shipment from the University of Virginia without proper declaration that they were hazardous. Dangerous Goods offered for shipment are highly regulated and must be shipped by trained personnel. An individual (faculty members, staff and students) who violates the requirements for shipping or transporting hazardous materials is liable for up to \$250,000 in fines and imprisonment for up to 5 years. It is therefore critical that you alert your faculty and everyone in your laboratories to (a) the need to follow hazardous shipping regulations and (b) to formally inform any contractors (e.g. movers or installers) entering a laboratory of the hazards in the lab. Environmental Health and Safety (EHS) is ready to assist lab personnel. If anyone plans on shipping hazardous materials (radioactive, biological or chemical which includes dry ice) they must have received "Dangerous Goods" training (available from EHS) or contact EHS for assistance.

Any laboratory that plans on moving equipment or hazardous materials should ask EHS for assistance to be sure that moving company employees are properly trained and that materials are properly packaged for transportation.

Dangerous Good	EHS Contact	Shipped by
Hazardous Chemicals ¹	<i>Adam Peters</i> (app5a@virginia.edu)	EHS personnel <u>only</u>
Dry Ice only ¹	<i>Adam Peters</i> (app5a@virginia.edu)	EHS personnel upon request
Radiological materials ¹	<i>Mike Cohen</i> (mlc6f@virginia.edu) or <i>Trevor Thomas</i> (tkt4j@virginia.edu)	EHS personnel <u>only</u>
Infectious Substances or other biological materials ²	<i>Erica Pearce</i> (epearce@virginia.edu)	Trained shipper in lab or department

¹All shipments of chemical or radiological materials MUST be shipped by EHS personnel. Additionally, unregulated materials on dry ice will also be shipped by EHS upon request. Materials will be picked up and taken back to the EHS office where they will be properly packaged, marked, labeled and documented for shipment. The lab will provide a PTAO to EHS to pay for shipping costs, which will be recovered from the lab of origin via cost transfer (a FedEx account number will be requested for radiological packages).

² For questions about the types of biological materials that are regulated or infectious substance training contact Erica Pearce. Training classes are provided on a monthly basis and schedules can be found at: <http://vprgsecure.web.virginia.edu/bio/training.cfm>

Please visit our website for updates and more information on Dangerous Goods at:

Shipments containing Dry Ice

13. **Especially hazardous chemicals also include:**

Picric acid, which is highly shock sensitive

Ethers, which can form highly unstable peroxides

Triethyl aluminum, which is highly pyrophoric

Lithium aluminum hydride, which is highly water reactive

Use extreme caution when handling and using all chemicals.

Common Solvents

Name	Freezing pt. F (C)	Boiling Pt. F (C)	Class
Acetone	-4 (-20)	133 (56)	IB
Acetonitrile	46 (6)	179 (82)	IB
Benzene	12 (-11)	176 (80)	IB
Butanol	84 (29)	243 (117)	IC
Carbon disulfide	-22 (-30)	115 (46)	IB
Cyclohexane	-4 (-20)	179 (82)	IB
Dioxane	52 (12)	214 (101)	IB
Ethanol	55 (13)	173 (78)	IB
Diethyl ether	-49 (-45)	95 (35)	IA
Heptane	25 (-4)	209 (98)	IB
Hexane	-7 (-22)	156 (69)	IB
Methanol	52 (11)	147 (64)	IB
Methyl ethyl ketone	16 (-9)	176 (80)	IB
Octane	56 (13)	258 (126)	IB
Pentane	-40	97 (36)	IA
2-Propanol	53 (12)	181 (83)	IB
THF	6 (-14)	151 (66)	IB
Toluene	40 (4)	231 (111)	IB
Xylene	81 (27)	281 (138)	IC

CHAPTER 4

COMPRESSED GASES AND GAS REGULATORS

1. *Know the contents of a cylinder and be familiar with the properties of that gas.* Never use a cylinder which can not be positively identified; cylinder color coding varies among gas vendors and is an unreliable identifier of cylinder contents.
2. **All cylinders in operation or not, must be firmly secured at all times by an adequate bench- or wall-mounted cylinder clamp or chain.** Keep in mind that breakage of the valve body on a compressed gas cylinder can easily transform the cylinder into a “torpedo” capable of going through concrete walls!
3. **When ordering new gas cylinders, consider purchasing gases in refillable containers.** The disposal costs of empty containers often offset the increased initial cost of having to purchase a larger amount of gas.
4. **Upon receipt of a new cylinder from the vendor immediately check the cylinder valve for leaks with a soap solution.** Leaks in cylinders should promptly be reported to the vendor for replacement.
5. **When installing a new cylinder, write your name on the cylinder information tag and attach it to the valve stem.**
6. **Use cylinders only with matched connectors and proper Compressed Gas Association (CGA) regulator.** Never install cylinder adapters on a regulator. Teflon tape must never be used on any CGA cylinder valve fitting.
7. **Oxygen regulators should be used only on oxygen tanks.** Contamination of oxygen regulators with the oil present in other gases can result in a serious explosion hazard when the regulator is again used for oxygen.
8. **Leak test all connections to a cylinder with a soap solution.** CAUTION! Any gas, regardless of its health hazard may cause asphyxiation by displacing oxygen.
9. **Pressure-relief devices protecting equipment attached to cylinders of flammable, toxic, or otherwise hazardous gases should be vented to an exhaust duct or fume hood.**
10. **When not in use, the regulators on cylinders should be depressurized.** If the cylinder is not to be used for a long time, the regulator must be removed. Never leave partly assembled apparatus attached to gas cylinders. Never attempt to refill a cylinder.
11. **When storing or moving a cylinder, always attach the iron safety hood securely to protect the valve stem, and transport gas cylinders of size 2 or larger only on a**

specifically designed wheeled cart.

12. **Cylinders should be located in the lab so that the cylinder valve is accessible at all times.** The main cylinder valve should be closed as soon as it is no longer necessary that it be open (i.e., it should never be left open when the equipment is unattended or not operating.) When storing or moving a cylinder, have the cap in place to protect the valve stem and never expose cylinders to temperatures higher than 50 Centigrade.
13. **Cylinders of compressed gases must be handled as high energy sources and therefore as potential explosives.** Cylinder valves should be opened slowly. Never tamper with any part of a valve such as the safety relief or packing nuts.
14. **A cylinder should never be emptied to a pressure lower than 172kPa (25 psig):** leave a slight pressure to keep contaminants out and notify the vendor with a note if draw-down occurs. Empty cylinders should not be refilled by anyone except the gas supplier. Remove the empty cylinder regulator and replace the valve cap. Keep the empty cylinder chained until pickup by the gas vendor. Be sure that a cylinder tag is attached and indicates the proper status of the cylinder (full, partially full, empty).
15. **Cylinder discharge lines should be equipped with approved check valves** to prevent inadvertent contamination of cylinders that are connected to a closed system where the possibility of flow reversal exists. Sucking back is particularly troublesome in the case of gases used as reactants in a closed system. If there is a possibility that a cylinder has been contaminated, it should be so labeled and returned to the supplier.
16. **When ordering toxic or flammable gases, whenever possible request a Flow Restrictor cylinder Valve.** The FRV orifice considerably reduces the full-open leak rate in event of a major leak (e.g., regulator diaphragm failure).
17. **Hydrogen and Oxygen cylinders must be kept 20 ft apart.**

CHAPTER 5

TOOL AND MACHINE SAFETY

Following a few basic safety procedures will assure a hazard-free experience for you and others working with tools in the laboratory.

1. Safety glasses (see Appendix 1) are required when using equipment in the shop area.
2. Remove rings, watches, bracelets, pendants and neckties which may be caught in moving machinery. Roll-up your long sleeves and secure long hair for the same reason.
3. Do not operate any shop equipment unless you have been instructed in its safe operation.
4. Report all injuries (and near misses), no matter how small, to the Safety Committee or the Department Chairman.
5. Always clean up the work areas before you leave.

APPENDIX 1

EYE PROTECTION

SAFETY GLASSES (see definition below) ARE THE MINIMUM EYE PROTECTION REQUIRED AND MUST BE WORN AT ALL TIMES IN ALL EYE PROTECTION AREAS. Side shields offer some protection from objects that approach from the side, but may not provide adequate protection from splashes. Goggles should be used when significant splash hazard exists, such as when handling concentrated acids and bases. Special eye protection is required for the handling of toxic chemicals, for welding, machining operations, etc.

Special protection is required when activities take place involving: (A) Corrosive or other chemically hazardous materials; (B) Hot molten metals; (C) Heat treatment; (D) Gas or electric arc welding; (E) Machine shop operations; (F) Vacuum evaporation, use of cryogenic apparatus or any evacuated experimental system where an implosion hazard exists; (G) Operation of high pressure reactors and reactions conducted in glass systems at any temperature or pressure.

Acceptable eye protection includes:

1. **Industrial safety glasses with side shields**, provided by you or obtained from your advisor.
2. **Your prescription glasses with plastic lenses or impact resistant glass lenses and lens-retaining frames with the addition of side shields**. Alternatively, safety glasses can be worn over prescription glasses.
3. **Visitor safety glasses with side shields that meet ANSI Z87.1-1979 standards**. These can be worn over prescription glasses and are acceptable for temporary use in laboratory areas. It is the host's responsibility to provide adequate eye protection for lab guests.

APPENDIX 2

DISPOSAL GUIDELINES

Department of Environmental Health & Safety
Chemical Waste Pickup: 982-4911

Waste Segregation We prefer to keep certain types of chemicals separated at the time of disposal. This method not only lowers disposal costs for the University, but also decreases the chances of incompatible materials from being added together. **Keep the following groups to themselves whenever possible.**

- 1 Non-halogenated organic solvents, <5% water
- 2 Non-halogenated organic solvents, >5% water
- 3 Halogenated solvents (% water unimportant)
- 4 Solutions containing compounds of the following metals: arsenic, barium, cadmium, chromium, lead, silver and selenium.
- 5 Any solution containing mercury or its compounds. (Mercury/mercury compounds should be kept separate from any liquid whenever possible.)
- 6 Acids, organic
- 7 Acids, mineral
- 8 Bases, organic
- 9 Bases, mineral
- 10 Acyl Halides (e.g. acetyl chloride, thionyl chloride, benzoyl chloride)
- 11 Cyanides
- 12 Sulfides
- 13 Organic peroxides
- 14 Inorganic Oxidizers
- 15 Photographic fixer
- 16 Photographic developer
- 17 Photographic stop bath
- 18 Water-reactive compounds (e.g. sodium, butyllithium, grignard reagents)
- 19 Pesticides
- 20 Oils
- 21 Paints
- 22 Formaldehyde Solutions

Do not put acidic or basic waste (pH <3 or >9) in metal cans. Metal cans corrode in a very short time. Keep acids and bases separated from hydrocarbons and ethers.

When possible, keep all carcinogens/mutagens separate from other waste. Keep aqueous wastes separate from organic solvents. Keep halogenated solvents and wastes separate from non-halogenated solvents.

Containers **Do not put hazardous waste down the sink or in the trash.** If you

And Labels are not sure if a chemical is hazardous, call EHS.

EHS provides plastic-coated, 1-gallon glass and DOT-approved 5-gallon carboys.

Chemically contaminated needles should be placed in Sharps-a-gator boxes, and will be disposed of by EHS.

All chemical waste must be deposited in properly labeled waste containers. According to the Virginia Department of Waste Management, each waste container **MUST** be marked with a hazardous waste sticker (Appendix 2A). Any containers issued by our office will already contain this sticker. If you plan to use your own bottles as waste receptacles, you can receive the required stickers by contacting our office, (2-4911). In addition to waste stickers, all waste containers **MUST** contain a waste disposal label (Appendix 2A) issued by EHS. This includes chemicals still in their original containers. Waste will not be picked up if it is not labeled properly. If you need new labels, let us know and we will deliver them on our pickup.

Both the label and its no-carbon-required copy should be affixed to the waste container by a single piece of tape across the top of the label, or in such a way that we can remove the copy when we pick up the waste.

Information that is absolutely required on the chemical waste label includes:

- a. The name of all possible contents, including stains, water, or any solvents.
Do not use abbreviations or formulas.
- b. The percentages of each component (total must equal 100%)
- c. The total quantity.
- d. The pH of the waste liquids if it is suspected to be below pH3 or above pH 10.
- e. Also include your name, date, department building and room number where the waste is located, phone number, and lab director.

Important: Disposal companies will not accept unknown chemicals. You must make every possible effort to accurately describe the contents of each container. This means tracking down and questioning previous lab occupants if necessary.

DO NOT FILL CONTAINERS TO THE TOP. Fill plastic carboys **ONLY** to the fill line. Leave about “2 at the top of all other containers. All waste must reside in closed, non-leaking containers. **Do Not** use flasks or test tubes with stoppers, beakers with parafilm, or bottles with ground glass stoppers. The outside of the waste container must be reasonable clean. Do Not put liquids (especially phenol) in bottles designed for solids.
They Leak!

The Virginia Department of Waste Management has stated that all chemical waste containers must remain **CLOSED** (capped) between chemical waste additions. When

chemical waste containers are left uncapped, laboratory personnel are risk of chemical exposure due to inhalation of chemical vapors.

Acidic solutions containing METALS (arsenic, barium, cadmium, chromium, lead, silver) should NOT go in 5-gallon carboys.

The Office of Environmental Health & Safety does not pick up empty bottles. They may be triple rinsed and discarded. They will supply empty bottles, as well as 5-gallon cans and carboys, for waste disposal. Call in advance for these items and we will bring them with your regular pickup.

Ethers tend to form extremely explosive compounds over time.

Therefore, date all ether cans. Do Not keep an open ether can for more than 1 month, or an unopened can for more than 4 months.

If you have an old ether can, label as waste call EHS for pick up.

Do not attempt to open bottles of DRY picric acid. This is an extreme explosion hazard!

Any dry bottles of picric acid should be labeled as waste, and picked up by EHS staff.

Other chemicals requiring special precautions include triethyl aluminum (which is highly pyrophoric) and lithium aluminum hydride (which is highly water reactive)

Do Not accumulate more than five 5-gallon cans or carboys, or more than ten gallons in bottles. Larger pickups will have to be scheduled separately.

Call 2-4911 or visit keats.admin.virginia.edu to schedule waste pick-up. Chemical waste will be picked up within three working days from the date it is called in.

APPENDIX 2A
STANDARD WASTE DISPOSAL LABELS

Hazardous Waste Label (SAMPLE ONLY – Obtain original from EHS)

<p style="text-align: center;">HAZARDOUS WASTE</p> <p style="text-align: center;">KEEP CONTAINERS CLOSED AT ALL TIMES DO NOT FILL WITHIN 2” OF CONTAINER TOP</p> <p>Hazardous contents:</p> <p>For emergencies or waste Pick Up, Call EHS 982-4911</p>

This label must be affixed to any waste container. Use the EHS-issued self-adhesive labels, which come in various sizes..

Waste Identification Label (SAMPLE ONLY – Obtain original from EHS)*(attach label here)*

U.Va. Office of Environmental Health And Safety

Chemical Safety Division

WASTE CHEMICAL/BIOHAZARD

IDENTIFICATION LABEL Date: _____

For Pick-up Call 982-4911

Lab Director: _____ Bldg: _____
 Department: _____ Rm. #: _____
 Your Name: _____ Phone #: _____

Waste Information: Please fill out *one label (with its copy) for each container* of waste. List *all* contents - including water. Indicate each component's percentage in the mixture (if unknown, approximate). Use proper chemical names. **DO NOT** use chemical formulas, structures or uncommon abbreviations. Please print. **DO NOT WRITE IN SHADED AREA.**

Waste Contents

Chemical Name	Percentage (%)

TOTAL PERCENTAGE (%) MUST EQUAL 100%

Total quantity in this container: _____

pH: _____

OFFICE USE ONLY

SA SA Before pH SEG:
Labeling

THIS MATERIAL IS POTENTIALLY HAZARDOUS. FEDERAL AND STATE LAW PROHIBITS IMPROPER DISPOSAL. IF FOUND, CONTACT THE OFFICE OF ENVIRONMENTAL HEALTH AND SAFETY, (804) 982-4911.

UPI-942215

APPENDIX 3

HAZARD REVIEW CHECKLIST
DEPARTMENT OF CHEMISTRY
UNIVERSITY OF VIRGINIA

The health and safety of you and your colleagues in the Department is your primary responsibility at all times. The experiment itself is secondary to safe lab practices.

This review checklist has been developed at the University of Delaware and is adapted from industrial hazard review forms in current use. The form is presented here with only minor modifications. Use it to review safety factors in your experimental equipment design and projected operating methods. The department safety manual and library safety references are good sources of design information. The completed form should be reviewed and approved by your advisor.

Review and approval to operate your equipment is not a blanket approval of safety status. The actual responsibility for safe operation is with the researcher.

Date: _____ Lab Location:

Title of Experiment:

Researcher: _____ Office Location:

Advisor:

Unusual Hazards: (Toxic gases, flammable solvents, flammable gases, high pressure gas or liquid, biological hazard, carcinogen, radiation hazard)

EMERGENCY SHUTDOWN PROCEDURE

Label all experimental equipment with emergency shutdown information so that a non-operator can easily shutdown your equipment. Information sheets and clipboards are available in the Department office.

Is Sequence important?

Device	Shutdown Location

Special First Aid Procedures:

Where is the nearest:

Evacuation Alarm
Exit

Fire Extinguisher
Safety Shower
Eyewash Station

Use the following checklist as a reminder to avoid unsafe practices and conditions in your equipment.

ELECTRICAL

Yes No N/A

- | | | |
|-------|----|--------------------------------------------------------------------------------|
| _ _ _ | 1) | Are power cords of adequate design, inspected, and in safe condition? |
| _ _ _ | 2) | Are voltages guarded? |
| _ _ _ | 3) | Have you considered static electricity hazards? |
| _ _ _ | 4) | Are switches labeled and accessible; i.e., not in potentially hazardous areas? |
| _ _ _ | 5) | Should electrical plugs and switches be explosion proof? |
| _ _ _ | 6) | Is over-temperature shutdown of heaters necessary, and if so, provided? |

- ___ 7) Is the test safe if electrical service is interrupted or fails?
- ___ 8) Are ground fault interrupters in place where needed?

MECHANICAL

- ___ 1) Are pinch points and exposed moving parts marked or guarded?
- ___ 2) Is the unit physically stable or mechanically anchored?
- ___ 3) Are cables, ropes, chainfalls, and/or pulleys the right size and have they been inspected and judged in safe condition?
- ___ 4) Is protection against backlash from cables, pulleys, or ropes provided if they break?
- ___ 5) Are proper lifting devices being used?
- ___ 6) Are mechanical shutdown interlocks provided if needed?
- ___ 7) Have rotating parts been checked for balance?

PRESSURE - PNEUMATIC HYDRAULIC AND STEAM

- ___ 1) Do gauges have blow-out backs and safety fronts, or alternately, read by mirror?
- ___ 2) Are relief ports and gauge blow-outs directed so that discharge does not constitute a hazard if they blow?
- ___ 3) Are adequate relief devices installed in proper locations? (No valves between device and source.)
- ___ 4) Are pressure ratings adequate? (Piping, fittings, vessels, valves, gauges, etc.)
- ___ 5) Do cylinder regulators have required inspections?
- ___ 6) Are cylinders properly secured?
- ___ 7) Do pressure vessels have current inspections?
- ___ 8) Are valves accessible; i.e., not in potentially hazardous areas? (Are valve stems of high pressure valves located above the operator's head or directed upwards?)
- ___ 9) Are flexible pressure lines secured to protect personnel in case of failure?
- ___ 10) Have safety relief valves been inspected and tested at set-point condition?
- ___ 11) Is nonmetallic tubing safe for this service? (Inert fluids, low pressure, temperature, static discharge.)

CHEMICAL

- ___ 1) Have you reviewed the Material Safety Data Sheets (MSDS) to determine hazards and handling procedures for test materials?
- ___ 2) Is an appropriate hazard warning tag posted outside your lab?
- ___ 3) Are materials of construction proper considering their recommended service as well as pH, chlorides, chemical contaminants, temperature, pressure, stress, cycling, and test duration?
- ___ 4) Are experiments placed in chemically resistant trays that will keep reagents from spreading in case of breakage?
- ___ 5) Have you eliminated all ignition sources near flammable chemicals (e.g., stirring motors, hot plates, powerstats, open flames, temp. baths, etc.)?
- ___ 6) Have you checked for hazardous reactions among chemicals in this test, chemicals

potentially present in a common exhaust or drain system, or chemicals stored nearby?

GENERAL

- ___ 1) Is test area free from tripping hazards and sharp edges?
- ___ 2) Are automatic shutdown devices required to protect personnel and equipment?
- ___ 3) Is the test safe if air, electricity, steam, or vacuum is interrupted or fails?
- ___ 4) Are area fire extinguisher proper type ("A", paper and wood; "B", oil solvent; "C", electrical) and are additional extinguisher needed?
- ___ 5) Is overhead clearance 7 feet? If not, is obstacle clearly marked?
- ___ 6) Are all containers labeled with contents, date, and person responsible?
- ___ 7) Are barricades and shields sufficient to prevent injury and protect equipment?
- ___ 8) Are signs and/or tags large enough and properly located to be easily seen?
- ___ 9) Does test require securing loose clothing and removing jewelry?
- ___ 10) Are inspection dates current on ladders, safety belts, or scaffolds required for overhead work?
- ___ 11) Have you planned an emergency escape route?
- ___ 12) Are good housekeeping practices being observed in the test area?
- ___ 13) Does noise level exceed 90 dB?
- ___ 14) Are personnel protected from hot/cold surfaces? (Steam lines, hot plates, etc.)
- ___ 15) Is special protective clothing, respirators, or first aid equipment provided and in good repair?
- ___ 16) Is dust level within allowable limits (10 mg/mL if nontoxic)?
- ___ 17) Will there be exposure of personnel to hazardous vapors?
- ___ 18) Is a special spill control procedure required?
- ___ 19) Is hood face air velocity adequate for the test being conducted?
- ___ 20) Is hood function impaired by air disturbances near the hood?

PERSONAL PROTECTIVE EQUIPMENT

FOR OPERATOR:

FOR VISITORS:

SPECIAL STANDBY:
(Emergency use)

GENERAL QUESTIONS

1. What are the experimental equipment limitations? (Temp., press., electrical, rpm, other.)
2. What human or unusual material or equipment failures could lead to an accident?
(Review test set-up carefully for hidden hazards.)

3. What unusual hazards will be involved in dismantling this experiment?
4. List the inventory of supplies (chemicals, reagents, solvents) you will maintain for this experiment. Where will you store these materials?
5. Explain your procedure to dispose of hazardous materials and used equipment?
6. What routine maintenance and routine safety inspections will you apply to your equipment to insure hazard-free service? Describe the records you will maintain.

APPENDIX 4

REFERENCES

Bretherick, L., Handbook of Reactive Chemical Hazards, 3rd ed., Butterworths, 1985

Compressed Gas Association, Inc., Handbook of Compressed Gases, 1981, 507 pp., VanNostrand Co., N.Y.

Hoffman, J. M. and D. C. Master, (ed), "Chemical Process Hazard Review", ACS Symposium Series 274, American Chemical Society, 1985

Manning, W. R. D., and Labrow, S., High Pressure Engineering, London, L. Hill Co., 1971, 369 pp.

National Research Council. Committee on Hazardous Substances in the Laboratory. 1981, Prudent Practices for Handling Hazardous Chemicals in Laboratories, National Academy Press, 289 pp., Washington, D.C.

NFPA 49, Hazardous Chemicals Data, 1975, National Fire Prevention Assoc., Boston, Mass.

Sax, N. I., 1979, Dangerous Properties of Industrial Materials, 6th ed., 1118 pp., VanNostrand Co., N.Y.

University of Virginia Chemistry Department Departure Safety Release Form

Prior to departure, each student (undergraduate and graduate), postdoctoral fellow or technician must verify and sign this form to release official graduation materials or final pay authorization.

I verify that I have cleaned the laboratory space that I have used.

I have disposed of all chemicals, solvents, and research materials properly, in accordance with EH & S, Departmental or lab-specific guidelines.

I have labeled all research products, reagents, and equipment accurately, and have stored them properly in authorized chemical storage facilities. Accurate and updated records of all such materials can be found in the laboratory notebooks that I have left with my advisor.

Laboratory Participant

Date

Advisor

Date

EMERGENCY NOTIFICATION FORM

DATE: _____

TITLE OF EXPERIMENT: _____

RESEARCHER: _____

ADVISOR: _____

WARNINGS:

SPECIAL EMERGENCY PROCEDURES:

IN CASE OF EMERGENCY CALL:

1) _____ Ph. No.

2) _____ Ph. No.

3) _____ Ph. No.

POTENTIAL HAZARDS: (Toxic gases, Flammable solvents, Flammable gases, High pressure gas, Biological hazard, Radiation Hazard, etc.)

Laboratory Start Certification Form

YOU MUST COMPLETE THIS FORM PRIOR TO USING LABORATORY FACILITIES IN THE DEPARTMENT. GIVE YOUR SIGNED FORM TO SUSAN MARSHALL FOR INCLUSION IN YOUR FILE.

I CERTIFY BY MY SIGNATURE BELOW THAT I HAVE READ THE CHEMISTRY DEPARTMENT SAFETY MANUAL, SEPTEMBER 2011 REVISION.

I HAVE LOCATED THE NEAREST SAFETY SHOWER AND EYE-WASH STATION FOR THE SPACES I WILL USE.

I HAVE LOCATED THE FIRE EXTINGUISHER AND FIRE ALARM NEAREST TO MY WORK SPACE.

I HAD A QUESTION ABOUT SAFETY PROCEDURES DESCRIBED IN THIS MANUAL.

YES

NO

MY QUESTIONS WERE ANSWERED BY THE SAFETY COMMITTEE OR BY MY ACADEMIC ADVISOR.

YES

NO

SIGNED

NAME (print)

DATE
